Boule Con-Diff

CONTROL

IVD

Control	Lot	
Tri-Level	32609-5K	2027-01-29
Low	32609-51	2027-01-29
Normal	32609-52	2027-02-02
High	32609-53	2027-02-02

Open vial stability 14 days [1]

SWELAB

Swelab Alfa/Alfa Plus		CONTROL / LOT	L / 32609-51		CONTROL / LOT	N / 32609-52		CONTROL / LOT	H / 32609-53	
		Mean [2]	± Range [3]		Mean [2]	± Range [3]		Mean [2]	± Range [3]	
RBC	10^{12}/L	2.24	0.12		4.12	0.18		5.00	0.22	
MCV	fL	74.1	7.0		84.2	7.0		92.2	7.0	
RDW	%	11.8	4.5		11.7	4.5		11.9	4.5	
Hct	%	16.6	3.0		34.7	3.5		46.1	4.0	
Plt	10^9/L	76	15		228	30		522	60	
MPV	fL	12.4	1.5		12.4	1.5		12.5	1.5	
WBC	10^9/L	3.3	0.3		8.4	0.6		21.1	1.8	
Hgb	g/dL	6.0	0.3		12.1	0.4		15.7	0.5	
MCH	pg	26.8	2.0		29.4	2.0		31.4	2.0	
MCHC	g/dL	36.1	4.0		34.9	4.0		34.1	4.0	
Lymph#	10^9/L	1.6	0.4		4.3	0.6		13.6	1.5	
Mid#	10^9/L	0.1	0.3		0.3	0.5		0.7	1.0	
Gran#	10^9/L	1.6	0.8		3.8	1.0		6.8	2.5	
Lymph	%	48.1	7.0		50.2	5.0		64.5	5.0	
Mid	%	3.3	5.0		4.0	5.0		3.5	5.0	
Gran	%	48.6	7.0		45.8	5.0		32.0	5.0	

	1	2	3
bg	Трайност на отворен флакон 14 дни (bg)	Средно	Обхват
hr	Stabilnost otvorene bočice 14 dani (hr)	Sredstvo	Raspon
cs	Stabilita otevřené ampulky 14 dny (cs)	Průměr	Rozsah
da	Åben rør stabilitet 14 dage (da)	Middelværdi	Område
et	Stabiilsus avatud viaalis 14 päevade (et)	Keskmine	Vahemik
fr	Stabilité en flacon ouvert 14 jour (fr)	Moyenne	Intervalle
de	Stabilität geöffneter Flaschen 14 tage (de)	Mittelwert	Bereich
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)	Μέση	Αναμενόμενο εύρος α
it	Stabilità della fiala aperta 14 giorni (it)	Media	Intervallo
lv	Atvērt flakonu stabilitāti 14 dienas (lv)	Nozīmē	Diapazons
lt	Stabilumas atidarius buteliuką 14 dienas (lt)	Vidurkis	Intervalas
no	Åpen rør/glass stabilitet 14 dager (no)	Middelverdi	Område
pl	Trwałość po otworzeniu fiolki 14 Liczba dni (pl)	Średnia	Zakres
pt	Estabilidade após abertura do frasco 14 dias (pt)	Média	Intervalo
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)	Medie	Interval
ru	Стабильность вскрытого флакона 14 дней (ru)	Среднее	диапазон
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)	Srednja vrednost	Opseg
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)	Priemer	Rozsah
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)	Povprečno	Območje
es	Estabilidad de la cápsula abierta 14 dias (es)	Media	Intervalo
sv	Hållbarhet för öppnad flaska 14 dagar (sv)	Medelvärde	Intervall
tr	Açık şişe stabilitesi 14 gün (tr)	Ortalama	Aralığı

Boule Medical AB
Fagerstagatan 7
SE-163 53 Spånga,
Sweden

25122-5

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:
1. In the Main Menu go to “Q/C Menu” “*Enter Con/Cal” “ Control/Calibrator Assay Value Input”*
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:
1. In the Main Menu go to “Quality Control” “Input Assays”
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.:	Swelab Alfa	Instr.:	Swelab Alfa	Instr.:	Swelab Alfa
Lot:	32609-51	Lot:	32609-52	Lot:	32609-53
Prod.:	Con-Diff Low	Prod.:	Con-Diff Norm	Prod.:	Con-Diff High
ID:	3260951+	ID:	3260952+	ID:	3260953+
Exp.:	29-Jan-2027	Exp.:	2-Feb-2027	Exp.:	2-Feb-2027

1: BoulevCBAV3 BoulevCBAV3 BoulevCBAV3

2: 3260951+ 3260952+ 3260953+

3: 27012900001153030000 27020200002153030000 27020200003153030000

4: 20014018000000000000 20014018000000000000 20014018000000000000

5: 22412741700600316630 41218842701210434735 50022922701570546140

6: 07615124152682036140 22830124152942034940 52260125153142034140

7: 03303016040010301608 08406043060030503810 21118136150071006825

8: 11845481700335048670 11745502500405045850 11945645500355032050

9: 11116347001196235229 11114488067844191053 11113204715844600529